

Sepapure protein G FPLC column

Short guide





Note: For your own safety, read the instructions and observe the warnings and safety information on the device and in the instructions. Keep the instructions for future reference.

KNAUER Technical Support: Phone: +49 30 809727-111 (9-17h, Central European Time)
Fax: +49 30 8015010
E-Mail: support@knauer.net
Languages: German, English

Publisher: KNAUER Wissenschaftliche Geräte GmbH
Hegauer Weg 38
14163 Berlin
Germany
Phone: +49 30 809727-0
Fax: +49 30 8015010
Internet: www.knauer.net
E-Mail: info@knauer.net

Version information: Article number: V6012
Version number: 1.0
Last update: 2019/01/29

The information in this document is subject to change without prior notice.
For latest version of the instructions, check our website:
www.knauer.net/en/Support/User-manuals.



Legal information This document contains confidential information and may not be reproduced without written content of KNAUER Wissenschaftliche Geräte GmbH.

© KNAUER Wissenschaftliche Geräte GmbH 2019
All rights reserved.

AZURA® is a registered trademark of
KNAUER Wissenschaftliche Geräte GmbH.

Table of Contents

1.	Specifications	1
1.1	Resin specifications.....	1
1.2	Buffers	2
2.	Preparing the column	2
3.	Sample preparation	2
4.	Antibody purification	2
5.	Column regeneration	3
6.	Column storage	3

1. Specifications

Definition Sepapure Protein G columns are designed for the purification of antibodies and antibody fragments via the Protein G - Fc region and Protein G - Fab region interactions. The columns are designed to be used with low-pressure FPLC-type automated purification systems and operated below 3 bar (0.3 MPa, 2.96 atm).

Hardware specifications

Column housing	Polypropylene
Frits	Polyethylene (nominal 20 µm porosity)
Fittings	10/32
Bed volume	1 and 5 mL

1.1 Resin specifications

Resin name	Sepapure Protein G FF
Base matrix material	4% cross-linked, beaded agarose
Mean bead diameter	100 µm
Ligand	Recombinant Protein G
Static antibody binding capacity	> 15 mg/mL human IgG
Flow rate	> 300 cm/hr
pH stability	2 - 10 (short term) 3 - 9 (long term)



Note: Before use, inspect the column for damage. If any damage is observed, do not use the column.



Note: Flow rates shown in this manual are for guidance only. Always ensure that system pressure is below the maximum for the column and resin.

1.2 Buffers

Sepapure columns are designed to be used with most aqueous phase chromatography buffers. A suggested buffer system is shown below, although other buffers may be used. Please check the resin specifications for further details.

Equilibration buffer	20 mM sodium phosphate, pH 7.0
Wash buffer	20 mM sodium phosphate, pH 7.0
Elution buffer	0.1 M glycine, pH 2.5 - 3.0

2. Preparing the column

Sepapure Protein G columns are supplied with 20% ethanol as the storage buffer. This must be removed prior to purification.

Process Remove the end-plugs and connect the column to the control system, taking care to avoid introduction of air into the system. Do not over-tighten fittings as this can strip the screw connections and lead to column leakage. Flush the column with 3 to 5 Column Volumes (CVs) of binding buffer at a flow rate of 1 to 2 CV/min to remove the storage buffer.

Equilibrate the column with 3 to 10 CVs of binding buffer at 1 CV/min to ensure that pH, conductivity and UV_{280} signals are stable.

Result The column is now ready for use.

3. Sample preparation

Sepapure Protein G columns should only be used with clarified (particle-free) samples. Most antibodies will bind to Protein G in the pH range of 7.0 - 7.4 and under physiological conditions (such as phosphate buffered saline), although some may exhibit better binding at slightly higher pH.



Note: Avoid excessively high concentrations of NaCl as whilst this may give higher binding capacity, it also increases binding of other proteins to the solid phase.

4. Antibody purification

Load the sample at a flow rate of approximately 1 CV/min. Some Protein G - antibody binding kinetics and sample viscosities will require the flow rate to be adjusted. To avoid overloading the column, the amount of sample that is applied to the column should not exceed the binding capacity of the solid phase to the target molecule.

Once the sample has been loaded, wash the column with binding buffer until the UV_{280} trace reaches baseline.

If a wash buffer is used, then also wash the column with 3 to 5 CVs of this reagent.

Elute the protein using the selected elution buffer. A stepwise or linear gradient may be used to determine the precise elution point of the target antibody. Antibodies can be eluted in a pH range starting from 5.5 and decreasing to as low as 2.5. Strongly bound antibodies may require a lower elution pH. Initially, all elution fractions should be collected for further analysis. Buffer exchange and/or desalting might be required following elution and we recommend Sepapure Desalting Columns (010X460SPZ and 020X460SPZ) for this purpose.

Acid-labile antibodies should be neutralised immediately on elution with 1 M Tris-HCl, pH 9.0 (10 - 20% v/v).

5. Column regeneration

To remove contaminants (lipids, non-specifically bound proteins, etc.) from the column following elution, wash with a further 3 to 5 CVs of elution buffer, followed by 3 to 5 CVs of equilibration buffer.

Some contaminants may not be fully removed by the above procedure, in which case a more stringent column regeneration/sanitization procedure should be used. To remove hydrophobic contaminants, use 1 CV of a non-ionic detergent (eg.: 0.1% Triton-X100) at 37°C at a flow rate of 1 CV/min immediately followed by 5 to 10 CVs of equilibration buffer.

Alternatively, the column can be washed in 70% ethanol for 12 hours followed by 5 to 10 CVs of equilibration buffer.

Sanitization of the column may be performed using a solution containing 2% chlorhexidine and 20% ethanol. Allow to stand for 6 hours before re-equilibrating with 3 to 5 CVs of equilibration buffer.

6. Column storage

Columns should be stored in PBS supplemented with 20% ethanol.



Note: Columns should be stored at +4°C.
Do not freeze!

Science Together



031 336 90 00 • www.scantecnordic.se

Latest KNAUER instructions online:
<https://www.knauer.net/en/Support/User-manuals>

KNAUER
Wissenschaftliche Geräte GmbH
Hegauer Weg 38
14163 Berlin

Phone: +49 30 809727-0
Fax: +49 30 8015010
E-Mail: info@knauer.net
Internet: www.knauer.net